



## CERTIFICATE OF ANALYSIS

**11-9005-12 - Tissue Marking Dye GREEN**  
**2 oz Avantik (MTO)**

Lot Number: 156625

Expiration Date: 2024-August-31

### COMPONENTS

Source: SOL-9005A/156744

| Test                        | Specifications             | Results |
|-----------------------------|----------------------------|---------|
| Color                       | Green                      | Green   |
| Viscosity @ 20 - 30 C (cps) | $\geq 380$ and $\leq 580$  | 489     |
| pH                          | $\geq 7.0$ and $\leq 10.0$ | 8.3     |

For In Vitro Use. May be suitable for other laboratory applications. Consult standard histology and pathology texts and operating procedures for details.

Quality Control

This report is electronically produced and is valid without signature.