



CERTIFICATE OF ANALYSIS

**11-9007-12 - Tissue Marking Dye BLUE 2
oz Avantik (MTO)**

Lot Number: 157389

Expiration Date: 2024-September-30

COMPONENTS

Source: SOL-9007A/157540

Test	Specifications	Results
Color	Blue	Blue
Viscosity @ 20 - 30 C (cps)	≥ 380 and ≤ 580	507
pH	≥ 7.0 and ≤ 10.0	9.4

For In Vitro Use. May be suitable for other laboratory applications. Consult standard histology and pathology texts and operating procedures for details.

Quality Control

This report is electronically produced and is valid without signature.