



## CERTIFICATE OF ANALYSIS

**11-9003-12 - Tissue Marking Dye  
ORANGE 2 oz Avantik (MTO)**

Lot Number: 157825

Expiration Date: 2024-September-30

### COMPONENTS

Source: SOL-9003A/157963

Test	Specifications	Results
Color	Orange	Orange
Viscosity @ 20 - 30 C (cps)	$\geq 380$ and $\leq 580$	450
pH	$\geq 7.0$ and $\leq 10.0$	8.2

For In Vitro Use. May be suitable for other laboratory applications. Consult standard histology and pathology texts and operating procedures for details.

Quality Control

This report is electronically produced and is valid without signature.