



## CERTIFICATE OF ANALYSIS

**11-5103-14 - Stain Gill 3 Hematoxylin**  
**Avantik 4x32 oz (MTO)**

Lot Number: 158769

Expiration Date: 2024-March-31

### COMPONENTS

Source: SOL-5103/158967

Test	Specifications	Results
Color	Purple-Red	Purple-Red
pH @ above 20 C	$\geq 2.05$ and $\leq 2.35$	2.25
Specific Gravity (g/cm3)	$\geq 0.98$ and $\leq 1.10$	1.06
UV/VIS Absorbance (A)	$\geq 1.03$ and $\leq 1.23$	1.09
UV/VIS Wavelength (nm)	$\geq 555$ and $\leq 560$	557
Functional Test	Pass/Fail	N/A

For In Vitro Use. May be suitable for other laboratory applications. Consult standard histology and pathology texts and operating procedures for details.

Quality Control

This report is electronically produced and is valid without signature.