



CERTIFICATE OF ANALYSIS

**11-9002-12 - Tissue Marking Dye RED 2
oz Avantik (MTO)**

Lot Number: 159204

Expiration Date: 2024-October-31

COMPONENTS

Source: SOL-9002A/158493

Test	Specifications	Results
Color	Red	Red
Viscosity @ 20 - 30 C (cps)	≥ 380 and ≤ 580	560
pH	≥ 7.0 and ≤ 10.0	10.0

For In Vitro Use. May be suitable for other laboratory applications. Consult standard histology and pathology texts and operating procedures for details.

Quality Control

This report is electronically produced and is valid without signature.