



CERTIFICATE OF ANALYSIS

11-9001-12 - Tissue Marking Dye BLACK
2 oz Avantik (MTO)

Lot Number: 160078

Expiration Date: 2024-October-31

COMPONENTS

Source: SOL-9001A/160246

Test	Specifications	Results
Color	Black	Black
Viscosity @ 20 - 30 C (cps)	≥ 380 and ≤ 580	508
pH	≥ 7.0 and ≤ 10.0	7.0

For In Vitro Use. May be suitable for other laboratory applications. Consult standard histology and pathology texts and operating procedures for details.

Quality Control

This report is electronically produced and is valid without signature.