



## CERTIFICATE OF ANALYSIS

**RS4364-A - Optik Hematoxylin Type 2,  
500mL(MTO)**

Lot Number: 161964

Expiration Date: 2024-May-31

### COMPONENTS

Source: SOL-BU0421/162068

Test	Specifications	Results
pH @ above 22C	Report	2.46

For In Vitro Use. May be suitable for other laboratory applications. Consult standard histology and pathology texts and operating procedures for details.

Quality Control

This report is electronically produced and is valid without signature.