



## CERTIFICATE OF ANALYSIS

RS4028-B - Gill's Hematoxylin 1, 1 Gallon  
(MTO)

Lot Number: 162939

Expiration Date: 2024-June-30

### COMPONENTS

Source: SOL-5101/163092

Test	Specifications	Results
Color	Purple	Purple
pH @ above 20 C	$\geq 2.35$ and $\leq 2.50$	2.40
Specific Gravity (g/cm <sup>3</sup> )	$\geq 0.98$ and $\leq 1.10$	1.05
UV/VIS Absorbance (A)	$\geq 0.48$ and $\leq 0.64$	0.48
UV/VIS Wavelength (nm)	$\geq 555$ and $\leq 560$	557
Functional Test	Pass/Fail	N/A

For In Vitro Use. May be suitable for other laboratory applications. Consult standard histology and pathology texts and operating procedures for details.

Quality Control

This report is electronically produced and is valid without signature.