



CERTIFICATE OF ANALYSIS

11-5103-24 - Stain Gill 3 Hematoxylin
Avantik 4x1 gal (MTO)

Lot Number: 163917

Expiration Date: 2024-June-30

COMPONENTS

Source: SOL-5103/163600

Test	Specifications	Results
Color	Purple-Red	Purple-Red
pH @ above 20 C	≥ 2.05 and ≤ 2.35	2.09
Specific Gravity (g/cm ³)	≥ 0.98 and ≤ 1.10	1.06
UV/VIS Absorbance (A)	≥ 1.03 and ≤ 1.23	1.10
UV/VIS Wavelength (nm)	≥ 555 and ≤ 560	555
Functional Test	Pass/Fail	Pass/Fail

For In Vitro Use. May be suitable for other laboratory applications. Consult standard histology and pathology texts and operating procedures for details.

Quality Control

This report is electronically produced and is valid without signature.