



CERTIFICATE OF ANALYSIS

**RS4605-B - Decalcifier- EDTA - 1 gallon
(MTO)**

Lot Number: 164906

Expiration Date: 2025-January-31

COMPONENTS

Source: SOL-BU0462/164996

Test	Specifications	Results
pH @ above 22C	≥ 7.0 and ≤ 7.4	7.4
Specific Gravity	Report	1.080

For In Vitro Use. May be suitable for other laboratory applications. Consult standard histology and pathology texts and operating procedures for details.

Quality Control

This report is electronically produced and is valid without signature.