



CERTIFICATE OF ANALYSIS

**11-9004-12 - Tissue Marking Dye
YELLOW 2 oz Avantik (MTO)**

Lot Number: 166232

Expiration Date: 2025-February-28

COMPONENTS

Source: SOL-9004A/166373

Test	Specifications	Results
Color	Yellow	Yellow
Viscosity @ 20 - 30 C (cps)	≥ 380 and ≤ 580	412
pH	≥ 7.0 and ≤ 10.0	9.6

For In Vitro Use. May be suitable for other laboratory applications. Consult standard histology and pathology texts and operating procedures for details.

Quality Control

This report is electronically produced and is valid without signature.