



## CERTIFICATE OF ANALYSIS

**HXB00242E - MASTERTECH BLUING  
SOLUTION /Gallon**

Lot Number: 167656

Expiration Date: 2025-February-28

### COMPONENTS

Source: SOL-HXB002/167771

Test	Specifications	Results
Visual Inspection	Clear, colorless solution	Clear, colorless solution
pH	$\geq 8.00$ and $\leq 9.00$	8.33

For In Vitro Use. May be suitable for other laboratory applications. Consult standard histology and pathology texts and operating procedures for details.

Quality Control

This report is electronically produced and is valid without signature.