



CERTIFICATE OF ANALYSIS

**RS4386-500 - Acid Alcohol, 1%, 500mL
(MTO)**

Lot Number: 169892

Expiration Date: 2026-March-31

COMPONENTS

Source: SOL-6800/170038

Test	Specifications	Results
Color	Clear	Clear
pH @ above 20 C	≥ 1.00 and ≤ 2.00	1.32
Alcohol (%)	≥ 69.0 and ≤ 71.0	70.0
Specific Gravity @ 20C (g/cm3)	≥ 0.883 and ≤ 0.900	0.890

For In Vitro Use. May be suitable for other laboratory applications. Consult standard histology and pathology texts and operating procedures for details.

Quality Control

This report is electronically produced and is valid without signature.