



## CERTIFICATE OF ANALYSIS

**11-9001-12 - Tissue Marking Dye BLACK**  
**2 oz Avantik (MTO)**

Lot Number: 170348

Expiration Date: 2025-March-31

### COMPONENTS

Source: SOL-9001A/170503

Test	Specifications	Results
Color	Black	Black
Viscosity @ 20 - 30 C (cps)	$\geq 380$ and $\leq 580$	488
pH	$\geq 7.0$ and $\leq 10.0$	7.1

For In Vitro Use. May be suitable for other laboratory applications. Consult standard histology and pathology texts and operating procedures for details.

Quality Control

This report is electronically produced and is valid without signature.