



CERTIFICATE OF ANALYSIS

RS4595-1GL - OG-6, 1 Gallon (MTO)

Lot Number: 171332

Expiration Date: 2024-October-31

COMPONENTS

Source: SOL-6006/171044

Test	Specifications	Results
Color	Orange	Orange
pH @ above 20 C	≥ 4.25 and ≤ 4.45	4.42
Specific Gravity (g/cm ³)	≥ 0.800 and ≤ 0.887	0.840
UV/VIS Absorbance (A)	≥ 0.48 and ≤ 0.63	0.48
UV/VIS Wavelength (nm)	≥ 475 and ≤ 480	477
Functional Test	Pass/Fail	N/A

For In Vitro Use. May be suitable for other laboratory applications. Consult standard histology and pathology texts and operating procedures for details.

Quality Control

This report is electronically produced and is valid without signature.