



## CERTIFICATE OF ANALYSIS

**RS4363-B - Optik Bluing Solution Type 1,  
RTU, 1 Gallon (MTO)**

Lot Number: 171849

Expiration Date: 2025-April-30

### COMPONENTS

Source: SOL-BU0398/172024

Test	Specifications	Results
Color	Match Standard	Light Blue
pH	$\geq 7.90$ and $\leq 8.10$	8.10

For In Vitro Use. May be suitable for other laboratory applications. Consult standard histology and pathology texts and operating procedures for details.

Quality Control

This report is electronically produced and is valid without signature.