



32 Commerce Street, Springfield, NJ 07081
888.392.8411 | www.avantik-us.com

October 18, 2021

CERTIFICATE OF ANALYSIS

Product Name: 10% Neutral Buffered Formalin, 90 mL Prefill

Description: Avantik 10% Neutral Buffered Formalin, 90ml prefills, 48/case

Product No.: GR2030-90

Lot No.: 26484

Date of Manufacture: 10/15/2021

Expiration Date: 2023-10-01

Materials:

| | <u>Part No.</u> | <u>Lot No.</u> | <u>Exp. Date</u> |
|-------------------|-----------------|------------------|------------------|
| Container: | ES144002 | D130821 | N/A |
| Lid: | ES144202 | F211221, F281221 | N/A |
| Chemical: | NBF-275G | 1J003 | 2023-10-01 |

QC Testing:

| LOT NUMBER: 1J003 | | | |
|-----------------------------------|----------------------------------|----------------|--------------------|
| PROPERTY | SPECIFICATION | RESULTS | PASS / FAIL |
| pH | 6.90-7.10 @ 25°C | 7.04 @ 25°C | PASS |
| Formaldehyde Concentration | 9.5-10.5% (v/v) | 10.31% (v/v) | PASS |
| Appearance | Clear with No Particulate | Clear | PASS |
| Labeling | Correct Lot Number and Exp. Date | Correct | PASS |

Country of Origin: United States of America

Latex Free: This product contains no latex

Created By:
James Osborn
Quality Associate