



32 Commerce Street, Springfield, NJ 07081  
888.392.8411 | [www.avantik-us.com](http://www.avantik-us.com)

October 19, 2021

### CERTIFICATE OF ANALYSIS

**Product Name:** 10% Neutral Buffered Formalin, 4x1 Gallon Case

**Description:** Avantik 10% Neutral Buffered Formalin, 4x1 Gallon Case

**Product No.:** RS4060

**Lot No.:** 26591

**Date of Manufacture:** 10/18/2021

**Expiration Date:** 2023-10-01

**Materials:**

	Part No.	Lot No.	Exp. Date
Container:	P1G-1Gal	846527, 846462	N/A
Lid:	CAP-1 Gal-RMS-38	202105608211	N/A
Chemical:	NBF-275G	1J005	2023-10-01

**QC Testing:**

LOT NUMBER: 1J005			
PROPERTY	SPECIFICATION	RESULTS	PASS / FAIL
pH	6.90-7.10 @ 25°C	7.02 @ 25°C	PASS
Formaldehyde Concentration	9.5-10.5% (v/v)	10.31% (v/v)	PASS
Appearance	Clear with No Particulate	Clear	PASS
Labeling	Correct Lot Number and Exp. Date	Correct	PASS

**Latex Free:** This product contains no latex

Created By:  
James Osborn  
Quality Associate