



32 Commerce Street, Springfield, NJ 07081  
888.392.8411 | [www.avantik-us.com](http://www.avantik-us.com)

December 02, 2021

### CERTIFICATE OF ANALYSIS

**Product Name:** 10% Neutral Buffered Formalin, 5 Gallon Cube

**Description:** Avantik 10% Neutral Buffered Formalin, 5 Gallon Cube

**Product No.:** RS4061

**Lot No.:** 26978

**Date of Manufacture:** 12/01/2021

**Expiration Date:** 2023-11-01

**Materials:**

	Part No.	Lot No.	Exp. Date
Cubitainer:	CDC6378	11/08/2021	N/A
Spigot:	COM4401	1498097	N/A
Chemical:	NBF-275G	1K013	2023-11-01

**QC Testing:**

LOT NUMBER: 1K013			
PROPERTY	SPECIFICATION	RESULTS	PASS / FAIL
pH	6.90-7.10 @ 25°C	6.94 @ 25°C	PASS
Formaldehyde Concentration	9.5-10.5% (v/v)	10.16% (v/v)	PASS
Appearance	Clear with No Particulate	Clear	PASS
Labeling	Correct Lot Number and Exp. Date	Correct	PASS

**Country of Origin:** United States of America

**Latex Free:** This product contains no latex

Created By:  
James Osborn  
Quality Associate