



32 Commerce Street, Springfield, NJ 07081  
888.392.8411 | [www.avantik-us.com](http://www.avantik-us.com)

12/07/2021

### CERTIFICATE OF ANALYSIS

**Product Name:** 10% Neutral Buffered Formalin, 40 mL Prefill

**Description:** Avantik 10% Neutral Buffered Formalin, 40ml prefills, 48/case

**Product No.:** GR2030-40

**Lot No.:** 27124

**Date of Manufacture:** 2021-12-07

**Expiration Date:** 2023-11-01

**Materials:**

	Part No.	Lot No.	Exp. Date
Container:	ES146002	H111221, H121221	N/A
Lid:	ES144202	I291221	N/A
Chemical:	NBF-275G	1K012	2023-11-01

**QC Testing:**

LOT NUMBER: 1K012			
PROPERTY	SPECIFICATION	RESULTS	PASS / FAIL
pH	6.90-7.10 @ 25°C	7.01 @ 25°C	PASS
Formaldehyde Concentration	9.5-10.5% (v/v)	10.31% (v/v)	PASS
Appearance	Clear with No Particulate	Clear	PASS
Labeling	Correct Lot Number and Exp. Date	Correct	PASS

**Country of Origin:** United States of America

**Latex Free:** This product contains no latex

Created By:  
James Osborn  
Quality Associate