



32 Commerce Street, Springfield, NJ 07081
888.392.8411 | www.avantik-us.com

December 21, 2021

CERTIFICATE OF ANALYSIS

Product Name: 10% Neutral Buffered Formalin, 60 mL Prefill

Description: Avantik 10% Neutral Buffered Formalin, 60ml prefills, 54/case

Product No.: GR2030-60

Lot No.: 27263

Date of Manufacture: 12/17/2021

Expiration Date: 2023-12-01

Materials:

	Part Number	Lot Number	Exp. Date
Container	ES145003	H130221, E170221, E180221	N/A
Lid	ES144202	D231221	N/A
Chemical	NBF-275G	1L001 1L006	2023-12-01 2023-12-01

QC Testing:

LOT NUMBER: 1L001			
PROPERTY	SPECIFICATION	RESULTS	PASS / FAIL
pH	6.90-7.10 @ 25°C	7.05 @ 25°C	PASS
Formaldehyde Concentration	9.5-10.5% (v/v)	10.31% (v/v)	PASS
Appearance	Clear with No Particulate	Clear	PASS
Labeling	Correct Lot Number and Exp. Date	Correct	PASS
LOT NUMBER: 1L006			
PROPERTY	SPECIFICATION	RESULTS	PASS / FAIL
pH	6.90-7.10 @ 25°C	7.04 @ 25°C	PASS
Formaldehyde Concentration	9.5-10.5% (v/v)	10.16% (v/v)	PASS
Appearance	Clear with No Particulate	Clear	PASS
Labeling	Correct Lot Number and Exp. Date	Correct	PASS

Country of Origin: United States of America

Latex Free: This product contains no latex

Created By:
James Osborn
Quality Associate