



32 Commerce Street, Springfield, NJ 07081
888.392.8411 | www.avantik-us.com

December 22, 2021

CERTIFICATE OF ANALYSIS

Product Name: 10% Neutral Buffered Formalin, 2.5 Gallon Cube

Description: Avantik 10% Neutral Buffered Formalin, 2.5 Gallon Cube

Product No.: RS4060-A

Lot No.: 27265

Date of: Manufacture 12/21/2021

Expiration Date: 2023-12-01

Materials:

	<u>Part No.</u>	<u>Lot No.</u>	<u>Exp. Date</u>
Cubitainer:	CDC5894	11/11/2021	N/A
Spigot:	COM4401	917426	N/A
Chemical:	NBF-275G	1L007	2023-12-01

QC Testing:

LOT NUMBER: 1L007			
PROPERTY	SPECIFICATION	RESULTS	PASS / FAIL
pH	6.90-7.10 @ 25°C	6.98 @ 25°C	PASS
Formaldehyde Concentration	9.5-10.5% (v/v)	10.31% (v/v)	PASS
Appearance	Clear with No Particulate	Clear	PASS
Labeling	Correct Lot Number and Exp. Date	Correct	PASS

Latex Free: This product contains no latex

Created By:
James Osborn
Quality Associate