



32 Commerce Street, Springfield, NJ 07081
888.392.8411 | www.avantik-us.com

December 23, 2021

CERTIFICATE OF ANALYSIS

Product Name: 10% Neutral Buffered Formalin, 120 mL Prefill

Description: Avantik 10% Neutral Buffered Formalin, 120ml prefills, 48/case

Product No.: GR2030-120

Lot No.: 27264

Date of Manufacture: 12/22/2021

Expiration Date: 2023-12-01

Materials:

| | Part No. | Lot No. | Exp. Date |
|------------|----------|------------------|------------|
| Container: | ES142002 | K090221, K200221 | N/A |
| Lid: | ES143202 | L011121 | N/A |
| Chemical: | NBF-275G | 1L006 | 2023-12-01 |
| | | 1L007 | 2023-12-01 |

QC Testing:

| LOT NUMBER: 1L006 | | | |
|----------------------------|----------------------------------|--------------|-------------|
| PROPERTY | SPECIFICATION | RESULTS | PASS / FAIL |
| pH | 6.90-7.10 @ 25°C | 7.04 @ 25°C | PASS |
| Formaldehyde Concentration | 9.5-10.5% (v/v) | 10.16% (v/v) | PASS |
| Appearance | Clear with No Particulate | Clear | PASS |
| Labeling | Correct Lot Number and Exp. Date | Correct | PASS |
| LOT NUMBER: 1L007 | | | |
| PROPERTY | SPECIFICATION | RESULTS | PASS / FAIL |
| pH | 6.90-7.10 @ 25°C | 6.98 @ 25°C | PASS |
| Formaldehyde Concentration | 9.5-10.5% (v/v) | 10.31% (v/v) | PASS |
| Appearance | Clear with No Particulate | Clear | PASS |
| Labeling | Correct Lot Number and Exp. Date | Correct | PASS |

Country of Origin: United States of America

Latex Free: This product contains no latex

Created By:
James Osborn
Quality Associate