



32 Commerce Street, Springfield, NJ 07081
888.392.8411 | www.avantik-us.com

January 10, 2022

CERTIFICATE OF ANALYSIS

Product Name: 10% Neutral Buffered Formalin, 90 mL Prefill

Description: Avantik 10% Neutral Buffered Formalin, 90ml prefills, 48/case

Product No.: GR2030-90

Lot No.: 27365

Date of Manufacture: 01/06/2022

Expiration Date: 2023-12-01

Materials:

	<u>Part No.</u>	<u>Lot No.</u>	<u>Exp. Date</u>
Container:	ES144002	T020221	N/A
Lid:	ES144202	J131221, J141221	N/A
Chemical:	NBF-275G	1L012	2023-12-01

QC Testing:

LOT NUMBER: 1L012			
PROPERTY	SPECIFICATION	RESULTS	PASS / FAIL
pH	6.90-7.10 @ 25°C	7.04 @ 25°C	PASS
Formaldehyde Concentration	9.5-10.5% (v/v)	9.85% (v/v)	PASS
Appearance	Clear with No Particulate	Clear	PASS
Labeling	Correct Lot Number and Exp. Date	Correct	PASS

Country of Origin: United States of America

Latex Free: This product contains no latex

Created By:
James Osborn
Quality Associate