



32 Commerce Street, Springfield, NJ 07081
888.392.8411 | www.avantik-us.com

01/18/2022

CERTIFICATE OF ANALYSIS

Product Name: 10% Neutral Buffered Formalin, 40 mL Prefill

Description: Avantik 10% Neutral Buffered Formalin, 40ml prefills, 48/case

Product No.: GR2030-40

Lot No.: 27361

Date of Manufacture: 2022-01-17

Expiration Date: 2024-01-01

Materials:

	Part No.	Lot No.	Exp. Date
Container:	ES146002	H151221	N/A
Lid:	ES144202	J121221, D231221, F261221	N/A
Chemical:	NBF-275G	2A001	2024-01-01

QC Testing:

LOT NUMBER: 2A001			
PROPERTY	SPECIFICATION	RESULTS	PASS / FAIL
pH	6.90-7.10 @ 25°C	6.98 @ 25°C	PASS
Formaldehyde Concentration	9.5-10.5% (v/v)	10.16% (v/v)	PASS
Appearance	Clear with No Particulate	Clear	PASS
Labeling	Correct Lot Number and Exp. Date	Correct	PASS

Country of Origin: United States of America

Latex Free: This product contains no latex

Created By:
James Osborn
Quality Associate