



32 Commerce Street, Springfield, NJ 07081  
888.392.8411 | [www.avantik-us.com](http://www.avantik-us.com)

January 25, 2022

### CERTIFICATE OF ANALYSIS

**Product Name:** 10% Neutral Buffered Formalin, 2.5 Gallon Cube

**Description:** Avantik 10% Neutral Buffered Formalin, 2.5 Gallon Cube

**Product No.:** RS4060-A

**Lot No.:** 27597

**Date of: Manufacture** 01/24/2022

**Expiration Date:** 2024-01-01

**Materials:**

	<u>Part No.</u>	<u>Lot No.</u>	<u>Exp. Date</u>
Cubitainer:	CDC5894	12/06/2021	N/A
Spigot:	COM4401	1348133	N/A
Chemical:	NBF-275G	2A007	2024-01-01

**QC Testing:**

LOT NUMBER: 2A007			
PROPERTY	SPECIFICATION	RESULTS	PASS / FAIL
pH	6.90-7.10 @ 25°C	6.98 @ 25°C	PASS
Formaldehyde Concentration	9.5-10.5% (v/v)	10.46% (v/v)	PASS
Appearance	Clear with No Particulate	Clear	PASS
Labeling	Correct Lot Number and Exp. Date	Correct	PASS

**Latex Free:** This product contains no latex

Created By:  
James Osborn  
Quality Associate