



19 Chapin Road - Building C, Pine Brook, NJ 07058
800.783.9424 | avantik-us.com

April 12, 2022

CERTIFICATE OF ANALYSIS

Product Name: 10% Neutral Buffered Formalin, 20 mL Prefill

Description: Avantik 10% Neutral Buffered Formalin, 20ml prefills, 128/case

Product No.: GR2030-20

Lot No.: 28658

Date of Manufacture: 2022-04-11

Expiration Date: 2024-03-01

Materials:

	Part Number	Lot Number	Exp. Date
Container	ES147002	C240222, C300222, C310222	N/A
Lid	ES147202	A080422, A090422	N/A
Chemical	NBF-275G	2C017	2024-03-01

QC Testing:

LOT NUMBER: 2C017			
PROPERTY	SPECIFICATION	RESULTS	PASS / FAIL
pH	6.90-7.10 @ 25°C	7.02 @ 25°C	PASS
Formaldehyde Concentration	9.5-10.5% (v/v)	10.46% (v/v)	PASS
Appearance	Clear with No Particulate	Clear	PASS
Labeling	Correct Lot Number and Exp. Date	Correct	PASS

Country of Origin: United States of America

Latex Free: This product contains no latex

Created by:
James Osborn
Quality Associate