



19 Chapin Road - Building C, Pine Brook, NJ 07058  
800.783.9424 | [avantik-us.com](http://avantik-us.com)

April 11, 2022

### CERTIFICATE OF ANALYSIS

**Product Name:** 10% Neutral Buffered Formalin, 20 mL Prefill

**Description:** Avantik 10% Neutral Buffered Formalin, 20ml prefills, Bulk, 32/pk 24 pk/case (768/CS)

**Product No.:** GR2102-20

**Lot No.:** 28583

**Date of Manufacture:** 2022-04-08

**Expiration Date:** 2024-03-01

**Materials:**

	Part Number	Lot Number	Exp. Date
Container	ES147002	C220222	N/A
Lid	ES147202	A090422, A100422	N/A
Chemical	NBF-275G	2C017	2024-03-01

**QC Testing:**

LOT NUMBER: 2C017			
PROPERTY	SPECIFICATION	RESULTS	PASS / FAIL
pH	6.90-7.10 @ 25°C	7.02 @ 25°C	PASS
Formaldehyde Concentration	9.5-10.5% (v/v)	10.46% (v/v)	PASS
Appearance	Clear with No Particulate	Clear	PASS
Labeling	Correct Lot Number and Exp. Date	Correct	PASS

**Country of Origin:** United States of America

**Latex Free:** This product contains no latex

Created By:  
James Osborn  
Quality Associate