



19 Chapin Road - Building C, Pine Brook, NJ 07058
800.783.9424 | avantik-us.com

April 15, 2022

CERTIFICATE OF ANALYSIS

Product Name: 10% Neutral Buffered Formalin, 5 Gallon Cube

Description: Avantik 10% Neutral Buffered Formalin, 5 Gallon Cube

Product No.: RS4061

Lot No.: 28588

Date of Manufacture: 04/14/2022

Expiration Date: 2024-04-01

Materials:

	Part No.	Lot No.	Exp. Date
Cubitainer:	CDC6378	03/08/2022, 03/09/2022	N/A
Spigot:	COM4401	1033808	N/A
Chemical:	NBF-275G	2D003 2D004	2024-04-01 2024-04-01

QC Testing:

LOT NUMBER: 2D003

PROPERTY	SPECIFICATION	RESULTS	PASS / FAIL
pH	6.90-7.10 @ 25°C	6.96 @ 25°C	PASS
Formaldehyde Concentration	9.5-10.5% (v/v)	10.16% (v/v)	PASS
Appearance	Clear with No Particulate	Clear	PASS
Labeling	Correct Lot Number and Exp. Date	Correct	PASS

LOT NUMBER: 2D004

PROPERTY	SPECIFICATION	RESULTS	PASS / FAIL
pH	6.90-7.10 @ 25°C	6.96 @ 25°C	PASS
Formaldehyde Concentration	9.5-10.5% (v/v)	10.46% (v/v)	PASS
Appearance	Clear with No Particulate	Clear	PASS
Labeling	Correct Lot Number and Exp. Date	Correct	PASS

Country of Origin: United States of America

Latex Free: This product contains no latex

Created By:
James Osborn
Quality Associate