



19 Chapin Road - Building C, Pine Brook, NJ 07058
800.783.9424 | avantik-us.com

June 08, 2022

CERTIFICATE OF ANALYSIS

Product Name: 10% Neutral Buffered Formalin, 4x1 Gallon Case

Description: Avantik 10% Neutral Buffered Formalin, 4x1 Gallon Case

Product No.: RS4060

Lot No.: 29332

Date of Manufacture: 06/06/2022

Expiration Date: 2024-05-01

Materials:

| | Part No. | Lot No. | Exp. Date |
|------------|------------------|--------------|------------|
| Container: | P1G-1Gal | 910490 | N/A |
| Lid: | CAP-1 Gal-RMS-38 | 202204634684 | N/A |
| Chemical: | NBF-275G | 2E009 | 2024-05-01 |

QC Testing:

| LOT NUMBER: 2E009 | | | |
|----------------------------|----------------------------------|--------------|-------------|
| PROPERTY | SPECIFICATION | RESULTS | PASS / FAIL |
| pH | 6.90-7.10 @ 25°C | 6.97 @ 25°C | PASS |
| Formaldehyde Concentration | 9.5-10.5% (v/v) | 10.16% (v/v) | PASS |
| Appearance | Clear with No Particulate | Clear | PASS |
| Labeling | Correct Lot Number and Exp. Date | Correct | PASS |

Latex Free: This product contains no latex

Created By:
James Osborn
Quality Associate