



19 Chapin Road - Building C, Pine Brook, NJ 07058
800.783.9424 | avantik-us.com

06/09/2022

Certificate of Analysis

Product No.: CUNZF-5-G

Description: Zinc Formalin, 5 gallon cube with spigot, Each

Lot No.: 2115

Date of Manufacture: 2022-04-25

Expiration Date: 2024-04-25

Materials:

	<u>Part No.</u>	<u>Lot No.</u>	<u>Exp. Date</u>
Chemical:	CUNZF	2115	2024-04-25

QC Testing:

10% FORMALIN LOT: 2115			
PROPERTY	SPECIFICATION	RESULTS	PASS / FAIL
pH	5.60-6.20 @ 25°C	5.7 @ 25°C	PASS
Formaldehyde Concentration	9.9-10.5% (v/v)	10.20% (v/v)	PASS
Appearance	Clear with No Particulate	Clear	PASS
Labeling	Correct Lot Number and Exp. Date	Correct	PASS

Country of Origin: United States of America

Latex Free: This product contains no latex