



19 Chapin Road - Building C, Pine Brook, NJ 07058  
800.783.9424 | [avantik-us.com](http://avantik-us.com)

June 14, 2022

### CERTIFICATE OF ANALYSIS

**Product Name:** 10% Neutral Buffered Formalin, 20 mL Prefill

**Description:** Avantik 10% Neutral Buffered Formalin, 20ml prefills, 128/case

**Product No.:** GR2030-20

**Lot No.:** 29331

**Date of Manufacture:** 2022-06-06

**Expiration Date:** 2024-05-01

**Materials:**

	Part Number	Lot Number	Exp. Date
Container	ES147002	D160222, D170222, D280222, D290222, E140222, E190222, A1012222, L120821	N/A
Lid	ES147202	D230422, D240422, D250422, D260422, E190422, E200422, E210422	N/A
Chemical	NBF-275G	2E009 2E010 2F001	2024-05-01 2024-05-01 2024-06-01

**QC Testing:**

LOT NUMBER: 2E009			
PROPERTY	SPECIFICATION	RESULTS	PASS / FAIL
pH	6.90-7.10 @ 25°C	6.97 @ 25°C	PASS
Formaldehyde Concentration	9.5-10.5% (v/v)	10.16% (v/v)	PASS
Appearance	Clear with No Particulate	Clear	PASS
Labeling	Correct Lot Number and Exp. Date	Correct	PASS
LOT NUMBER: 2E010			
PROPERTY	SPECIFICATION	RESULTS	PASS / FAIL
pH	6.90-7.10 @ 25°C	7.09 @ 25°C	PASS
Formaldehyde Concentration	9.5-10.5% (v/v)	10.31% (v/v)	PASS
Appearance	Clear with No Particulate	Clear	PASS
Labeling	Correct Lot Number and Exp. Date	Correct	PASS



32 Commerce Street, Springfield, NJ 07081  
888.392.8411 | [www.avantik-us.com](http://www.avantik-us.com)

LOT NUMBER: 2F001			
PROPERTY	SPECIFICATION	RESULTS	PASS / FAIL
pH	6.90-7.10 @ 25°C	7.06 @ 25°C	PASS
Formaldehyde Concentration	9.5-10.5% (v/v)	10.31% (v/v)	PASS
Appearance	Clear with No Particulate	Clear	PASS
Labeling	Correct Lot Number and Exp. Date	Correct	PASS

**Country of Origin:** United States of America

**Latex Free:** This product contains no latex

Created by:  
James Osborn  
Quality Associate