



19 Chapin Road - Building C, Pine Brook, NJ 07058
800.783.9424 | avantik-us.com

August 17, 2022

CERTIFICATE OF ANALYSIS

Product Name: 10% Neutral Buffered Formalin, 20 mL Prefill

Description: Avantik 10% Neutral Buffered Formalin, 20ml prefills, 128/case

Product No.: GR2030-20

Lot No.: 30016

Date of Manufacture: 2022-08-05

Expiration Date: 2024-08-01

Materials:

	Part Number	Lot Number	Exp. Date
Container	ES147002	C250222, C260222, C270222, D270222, D280222, L110821, G100222, G150222, G160222	N/A
Lid	ES147202	E090422, E100422, E110422, F250422, F260422, G170422, G180422, G190422	N/A
Chemical	NBF-275G	2H001 2H002 2H004	2024-08-01 2024-08-01 2024-08-01

QC Testing:

LOT NUMBER: 2H001			
PROPERTY	SPECIFICATION	RESULTS	PASS / FAIL
pH	6.90-7.10 @ 25°C	7.04 @ 25°C	PASS
Formaldehyde Concentration	9.5-10.5% (v/v)	10.31% (v/v)	PASS
Appearance	Clear with No Particulate	Clear	PASS
Labeling	Correct Lot Number and Exp. Date	Correct	PASS
LOT NUMBER: 2H002			
PROPERTY	SPECIFICATION	RESULTS	PASS / FAIL
pH	6.90-7.10 @ 25°C	7.06 @ 25°C	PASS
Formaldehyde Concentration	9.5-10.5% (v/v)	10.46% (v/v)	PASS
Appearance	Clear with No Particulate	Clear	PASS
Labeling	Correct Lot Number and Exp. Date	Correct	PASS



[19 Chapin Road - Building C, Pine Brook, NJ 07058](#)
[800.783.9424 | avantik-us.com](#)

LOT NUMBER: 2H004			
PROPERTY	SPECIFICATION	RESULTS	PASS / FAIL
pH	6.90-7.10 @ 25°C	7.01 @ 25°C	PASS
Formaldehyde Concentration	9.5-10.5% (v/v)	10.31% (v/v)	PASS
Appearance	Clear with No Particulate	Clear	PASS
Labeling	Correct Lot Number and Exp. Date	Correct	PASS

Country of Origin: United States of America

Latex Free: This product contains no latex