



19 Chapin Road - Building C, Pine Brook, NJ 07058  
800.783.9424 | [avantik-us.com](http://avantik-us.com)

08/17/2022

### CERTIFICATE OF ANALYSIS

**Product Name:** 10% Neutral Buffered Formalin, 40 mL Prefill

**Description:** Avantik 10% Neutral Buffered Formalin, 40ml prefills, 48/case

**Product No.:** GR2030-40

**Lot No.:** 30017

**Date of Manufacture:** 2022-08-16

**Expiration Date:** 2024-08-01

**Materials:**

	Part No.	Lot No.	Exp. Date
Container:	ES146002	B121222, B131222	N/A
Lid:	ES144202	F261222, F271222	N/A
Chemical:	NBF-275G	2H002	2024-08-01

**QC Testing:**

LOT NUMBER: 2H002			
PROPERTY	SPECIFICATION	RESULTS	PASS / FAIL
pH	6.90-7.10 @ 25°C	7.06 @ 25°C	PASS
Formaldehyde Concentration	9.5-10.5% (v/v)	10.46% (v/v)	PASS
Appearance	Clear with No Particulate	Clear	PASS
Labeling	Correct Lot Number and Exp. Date	Correct	PASS

**Country of Origin:** United States of America

**Latex Free:** This product contains no latex