



19 Chapin Road - Building C, Pine Brook, NJ 07058
800.783.9424 | avantik-us.com

November 30, 2022

CERTIFICATE OF ANALYSIS

Product Name: 10% Neutral Buffered Formalin, 4x1 Gallon Case

Description: Avantik 10% Neutral Buffered Formalin, 4x1 Gallon Case

Product No.: RS4060

Lot No.: 31061

Date of Manufacture: 2022-11-18

Expiration Date: 2024-11-01

Materials:

	Part No.	Lot No.	Exp. Date
Container:	P1G-1Gal	221011-13400	N/A
Lid:	CAP-1 Gal-RMS-38	202208642107	N/A
Chemical:	NBF-275G	2K004	2024-11-01
		2K006	2024-11-01
		2K008	2024-11-01

QC Testing:

LOT NUMBER: 2K004			
PROPERTY	SPECIFICATION	RESULTS	PASS / FAIL
pH	6.90-7.10 @ 25°C	7.01 @ 25°C	PASS
Formaldehyde Concentration	9.5-10.5% (v/v)	10.31% (v/v)	PASS
Appearance	Clear with No Particulate	Clear	PASS
Labeling	Correct Lot Number and Exp. Date	Correct	PASS
LOT NUMBER: 2K006			
PROPERTY	SPECIFICATION	RESULTS	PASS / FAIL
pH	6.90-7.10 @ 25°C	6.97 @ 25°C	PASS
Formaldehyde Concentration	9.5-10.5% (v/v)	10.00% (v/v)	PASS
Appearance	Clear with No Particulate	Clear	PASS
Labeling	Correct Lot Number and Exp. Date	Correct	PASS
LOT NUMBER: 2K008			
PROPERTY	SPECIFICATION	RESULTS	PASS / FAIL
pH	6.90-7.10 @ 25°C	7.07 @ 25°C	PASS
Formaldehyde Concentration	9.5-10.5% (v/v)	10.31% (v/v)	PASS
Appearance	Clear with No Particulate	Clear	PASS
Labeling	Correct Lot Number and Exp. Date	Correct	PASS

Latex Free: This product contains no latex