

**Certificate of Analysis** 

1/3/2023

Product No.: RS4061

**Description:** Avantik Labeled 10% Neutral Buffered Formalin, 5 gallon cube

Lot No.: 31433

**Date of Manufacturing:** 12/27/2022

Expiration Date: 12/1/2024

10% Formalin Lot #: 2L009			
PROPERTY	<b>TEST SPECIFICATION</b>	RESULTS	
РН	6.90-7.10 @ 25°C	7.02@ 25°C	
Formaldehyde Concentration	9.5-10.5% (v/v)	10.31%(v/v)	
Appearance	Clear, colorless, no observable particulate	Conforms	

10% Formalin Lot #: 2L010			
PROPERTY	<b>TEST SPECIFICATION</b>	RESULTS	
PH	6.90-7.10 @ 25°C	6.95@ 25°C	
Formaldehyde Concentration	9.5-10.5% (v/v)	10.46%(v/v)	
Appearance	Clear, colorless, no observable particulate	Conforms	

10% Formalin Lot #: 2L011			
PROPERTY	<b>TEST SPECIFICATION</b>	RESULTS	
РН	6.90-7.10 @ 25°C	6.9@ 25°C	
Formaldehyde Concentration	9.5-10.5% (v/v)	10.31%(v/v)	
Appearance	Clear, colorless, no observable particulate	Conforms	

Country of Origin: United States of America Latex Free: This product contains no latex