



19 Chapin Road - Building C, Pine Brook, NJ 07058
800.783.9424 | avantik-us.com

CERTIFICATE OF ANALYSIS

**GR2056-BK - Tissue Marking Dye,
Avantik, Black, 2oz, 12/cs (MTO)**

Lot Number: 186724

Expiration Date: 2025-December-31

COMPONENTS

Source: SOL-9001A/187445

Test	Specifications	Results
Color	Black	Black
Viscosity @ 20 - 30 C (cps)	≥ 380 and ≤ 580	436
pH	≥ 7.0 and ≤ 10.0	8.7

For In Vitro Use. May be suitable for other laboratory applications. Consult standard histology and pathology texts and operating procedures for details.

Quality Control

This report is electronically produced and is valid without signature.