



19 Chapin Road - Building C, Pine Brook, NJ 07058
800.783.9424 | avantik-us.com

CERTIFICATE OF ANALYSIS

GR2056-BK - Tissue Marking Dye,
Avantik, Black, 2oz, 12/cs (MTO)

Lot Number: 191219

Expiration Date: 2026-February-28

COMPONENTS

Source: SOL-9001A/191365

Test	Specifications	Results
Color	Black	Black
Viscosity @ 20 - 30 C (cps)	>= 380 and <= 580	519
pH	>= 7.0 and <= 10.0	8.1

For In Vitro Use. May be suitable for other laboratory applications. Consult standard histology and pathology texts and operating procedures for details.

Quality Control
This report is electronically produced and is valid without signature.