



## CERTIFICATE OF ANALYSIS

**11-9007-12CS - Tissue Marking Dye,  
Avantik, Blue, 2oz, 12/cs (MTO)**

Lot Number: 176963

Expiration Date: 2025-August-31

### COMPONENTS

Source: SOL-9007A/177072

Test	Specifications	Results
Color	Blue	Blue
Viscosity @ 20 - 30 C (cps)	$\geq 380$ and $\leq 580$	440
pH	$\geq 7.0$ and $\leq 10.0$	8.2

For In Vitro Use. May be suitable for other laboratory applications. Consult standard histology and pathology texts and operating procedures for details.

Quality Control

This report is electronically produced and is valid without signature.