



19 Chapin Road - Building C, Pine Brook, NJ 07058
800.783.9424 | avantik-us.com

CERTIFICATE OF ANALYSIS

**GR2056-BL - Tissue Marking Dye,
Avantik, Blue, 2oz, 12/cs (MTO)**

Lot Number: 191223

Expiration Date: 2026-February-28

COMPONENTS

Source: SOL-9007A/191891

Test	Specifications	Results
Color	Blue	Blue
Viscosity @ 20 - 30 C (cps)	≥ 380 and ≤ 580	421
pH	≥ 7.0 and ≤ 10.0	8.0

For In Vitro Use. May be suitable for other laboratory applications. Consult standard histology and pathology texts and operating procedures for details.

Quality Control

This report is electronically produced and is valid without signature.