



19 Chapin Road - Building C, Pine Brook, NJ 07058  
800.783.9424 | avantik-us.com

## CERTIFICATE OF ANALYSIS

**GR2056-GN - Tissue Marking Dye,  
Avantik, Green, 2oz, 12/cs (MTO)**

Lot Number: 191221

Expiration Date: 2026-February-28

### COMPONENTS

Source: SOL-9005A/191367

Test	Specifications	Results
Color	Green	Green
Viscosity @ 20 - 30 C (cps)	$\geq 380$ and $\leq 580$	580
pH	$\geq 7.0$ and $\leq 10.0$	8.3

For In Vitro Use. May be suitable for other laboratory applications. Consult standard histology and pathology texts and operating procedures for details.

Quality Control

This report is electronically produced and is valid without signature.