



19 Chapin Road - Building C, Pine Brook, NJ 07058
800.783.9424 | avantik-us.com

CERTIFICATE OF ANALYSIS

**GR2056-OR - Tissue Marking Dye,
Avantik, Orange, 2oz, 12/cs (MTO)**

Lot Number: 186726

Expiration Date: 2025-November-30

COMPONENTS

Source: SOL-9003A/185971

Test	Specifications	Results
Color	Orange	Orange
Viscosity @ 20 - 30 C (cps)	≥ 380 and ≤ 580	405
pH	≥ 7.0 and ≤ 10.0	8.5

For In Vitro Use. May be suitable for other laboratory applications. Consult standard histology and pathology texts and operating procedures for details.

Quality Control

This report is electronically produced and is valid without signature.