



CERTIFICATE OF ANALYSIS

**11-9002-12CS - Tissue Marking Dye,
Avantik, Red, 2oz, 12/cs (MTO)**

Lot Number: 176533

Expiration Date: 2025-June-30

COMPONENTS

Source: SOL-9002A/176680

| Test | Specifications | Results |
|-----------------------------|----------------------------|---------|
| Color | Red | Red |
| Viscosity @ 20 - 30 C (cps) | ≥ 380 and ≤ 580 | 412 |
| pH | ≥ 7.0 and ≤ 10.0 | 10.0 |

For In Vitro Use. May be suitable for other laboratory applications. Consult standard histology and pathology texts and operating procedures for details.

Quality Control

This report is electronically produced and is valid without signature.