



CERTIFICATE OF ANALYSIS

11-5502-14 - Eosin Y 4x32 oz Avantik
RS4017-A (MTO)

Lot Number: 141871

Expiration Date: 2024-January-31

COMPONENTS

Source: SOL-5502/141873

Test	Specifications	Results
Color	Orange	Orange
pH @ above 20 C	≥ 4.42 and ≤ 4.57	4.48
Specific Gravity (g/cm ³)	≥ 0.890 and ≤ 0.910	0.910
UV/VIS Absorbance (A)	≥ 0.90 and ≤ 1.10	0.99
UV/VIS Wavelength (nm)	≥ 510 and ≤ 520	517
Functional Test	Pass/Fail	N/A

For In Vitro Use. May be suitable for other laboratory applications. Consult standard histology and pathology texts and operating procedures for details.

Quality Control

This report is electronically produced and is valid without signature.