



19 Chapin Road - Building C, Pine Brook, NJ 07058
800.783.9424 | avantik-us.com

CERTIFICATE OF ANALYSIS

Lot Number: 185003

RS4017-B - Eosin Y 4x1 Gal Avantik
RS4017-B (MTO)

Expiration Date: 2025-November-30

COMPONENTS

Source: SOL-5502/185123

| Test | Specifications | Results |
|---------------------------------------|-------------------------------|---------|
| Color | Orange | Orange |
| pH @ above 20 C | ≥ 4.42 and ≤ 4.57 | 4.42 |
| Specific Gravity (g/cm ³) | ≥ 0.890 and ≤ 0.910 | 0.910 |
| UV/VIS Absorbance (A) | ≥ 0.90 and ≤ 1.10 | 1.10 |
| UV/VIS Wavelength (nm) | ≥ 510 and ≤ 520 | 515 |
| Functional Test | Pass/Fail | N/A |

For In Vitro Use. May be suitable for other laboratory applications. Consult standard histology and pathology texts and operating procedures for details.

Quality Control

This report is electronically produced and is valid without signature.