



CERTIFICATE OF ANALYSIS

**RS4020-B - Harris Hematoxylin
(Advanced), GAL (MTO)**

Lot Number: 175654

Expiration Date: 2024-June-30

COMPONENTS

Source: SOL-BU0394/175779

Test	Specifications	Results
pH	≥ 2.50 and ≤ 2.75	2.61

For In Vitro Use. May be suitable for other laboratory applications. Consult standard histology and pathology texts and operating procedures for details.

Quality Control

This report is electronically produced and is valid without signature.