



19 Chapin Road - Building C, Pine Brook, NJ 07058
800.783.9424 | avantik-us.com

CERTIFICATE OF ANALYSIS

RS4021-A - Stain Gill 3 Hematoxylin Avantik 4x32 oz (MTO)

Lot Number: 179191

Expiration Date: 2025-April-30

COMPONENTS

Test	Specifications	Results
Color	Purple-Red	Purple-Red
pH @ above 20 C	≥ 2.05 and ≤ 2.35	2.22
Specific Gravity (g/cm3)	≥ 0.98 and ≤ 1.10	1.06
UV/VIS Absorbance (A)	≥ 1.03 and ≤ 1.23	1.21
UV/VIS Wavelength (nm)	≥ 555 and ≤ 560	555

For In Vitro Use. May be suitable for other laboratory applications. Consult standard histology and pathology texts and operating procedures for details.

Quality Control - This report is electronically produced and is valid without signature.