



19 Chapin Road - Building C, Pine Brook, NJ 07058  
800.783.9424 | avantik-us.com

## CERTIFICATE OF ANALYSIS

**RS4025-B - Hematoxylin Gills II, 1 Gal(MTO)**

Lot Number: 180559

Expiration Date: 2025-March-31

### COMPONENTS

Test	Specifications	Results
Color	Purple-Red	Purple-Red
pH @ above 20 C	$\geq 2.25$ and $\leq 2.45$	2.33
Specific Gravity (g/cm3)	$\geq 0.98$ and $\leq 1.10$	1.05
UV/VIS Absorbance (A)	$\geq 0.84$ and $\leq 1.00$	0.94
UV/VIS Wavelength (nm)	$\geq 555$ and $\leq 560$	558

For In Vitro Use. May be suitable for other laboratory applications. Consult standard histology and pathology texts and operating procedures for details.

Quality Control - This report is electronically produced and is valid without signature.