



19 Chapin Road - Building C, Pine Brook, NJ 07058  
800.783.9424 | avantik-us.com

## Certificate of Analysis

### Scotts Tap Water (RS4044)

Product Code: **RS4044**

Manufacture Date: January 23, 2024

Lot Number: **44010304**

Expiration Date: January 23, 2026

Test	Specification	Result
Appearance (clarity)	clear solution	clear solution
Appearance (color)	colorless	colorless
Color (APHA, Pt/Co)	< 5	<5
NaCl as TDS (100 x ppm)	42 - 48ppm	44ppm
pH @ 25°C	8.3 - 8.7	8.5

**Intended Use** - Product is intended for use in manufacturing procedures and laboratory procedures and protocols.

**Storage Information** - Unless noted on the product label, store the product under normal lab conditions in its tightly closed, original container. Do not pipet directly from the container or return unused portions to the container.

**Instructions for Handling and Use** - Please refer to the associated product label and Safety Data Sheet (SDS) for information regarding safety and handling of this product.

**Preparation** - All products are manufactured and tested according to established, documented procedures and methodology. Production documentation records manufacturing data, raw material traceability and testing history on a per lot basis. Balances, thermometers, and glassware are calibrated before first use and on a regular schedule with references traceable to NIST standards.