



19 Chapin Road - Building C, Pine Brook, NJ 07058
800.783.9424 | avantik-us.com

CERTIFICATE of ANALYSIS

Product: **Statlab Avantik UltraClear Xylene Substitute**
Product Code: **RS4058-A**

Log No: 41925
Date Rec'd/Test: 7/21/2023
Batch No.: 231289U

Test Description	Test Method	UOM	Specification / Min.--Max.		Test Results
Appearance	Visual	-	B,C,FFSM		PASS
Refractive index @ 20 °C	ASTM D-1218	-	1.4139	1.4209	1.4180
Specific gravity @ 15.6 °C/15.6 °C	ASTM D-4052	-	0.7430	0.7550	0.7482
Density @ 15.6 °C	SG x 8.32823	lb/gal	6.188	6.288	6.231
Water content	ASTM D-6304	% wt		0.0500	0.0101
Color, Saybolt	ASTM D-156	SCU	30		>30

***Shelf life is 24 months with container being stored in a cool, dry environment without moisture.**

Please consult the SDS for further information.

Univar Solutions represents only that the Product shall meet the specifications herein. All transactions involving this Product are subject to Univar Solutions' standard Terms and Conditions, available at www.univarsolutions.com or upon request. Univar Solutions makes no additional representations or warranties, express or implied, as to the Product.
www.univarsolutions.com

QA 6.20 x 1; 03/05/2019